



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Quality Instruments Services de México, S.A. de C.V.

***Fray Nicolás de Zamora No.73, Col. El Pueblito
Corregidora, Querétaro, México. C.P. 76900***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited
in accordance with the recognized International Standard:*

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the
operation of a laboratory quality management system
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Time and Frequency, Mechanical, Thermodynamic, Chemical and Optical
Calibration***
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

January 28, 2020

Issue Date:

January 07, 2022

Expiration Date:

February 28, 2024

Accreditation No.:

89224

Certificate No.:

L22-21

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a
continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjlabs.com*



Certificate of Accreditation: Supplement

Quality Instruments Services de México, S.A. de C.V.

Fray Nicolás de Zamora No.73, Col. El Pueblito

Corregidora, Querétaro, México. C.P. 76900

Contact Name: Alvaro Fernando Gomez Prieto Phone: 442-225-4758

Accreditation is granted to the facility to perform the following calibrations:

Time and Frequency

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Stopwatch ^o	60 s to 86 400 s	0.82 s/day	Direct Comparison Stopwatch Control Company CENAM Technical Guide

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Pressure Meter ^o	5 psi to 50 psi	0.1 psi	Digital Dwyer DPG-104 Manometer CENAM Technical Guide

Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Temperature Generation: Ovens, Furnaces, Muffles, Freezers, Incubators ^o	-40 °C to 1 000 °C	0.4 °C	Thermometer Extech Mod. 421502 CENAM Technical Guide
Relative Humidity Meter ^o	11.3 % RH to 97.4 % RH	2 % RH	Thermohygrometer Rotronic Mod. Hydroclip HK-25 CENAM Technical Guide

Chemical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
pH Meter ^o	4 pH	0.021 pH	Traceable, Standards Fermont CENAM Technical Guide
	7 pH	0.02 pH	
	10 pH	0.032 pH	
Conductivity Meter ^o	84 μ S/cm	1 μ S/cm	NIST Traceable Standards Hanna Solutions CENAM Technical Guide
	1 413 μ S/cm	5 μ S/cm	



Certificate of Accreditation: Supplement

Quality Instruments Services de México, S.A. de C.V.

Fray Nicolás de Zamora No.73, Col. El Pueblito

Corregidora, Querétaro, México. C.P. 76900

Contact Name: Alvaro Fernando Gomez Prieto Phone: 442-225-4758

Accreditation is granted to the facility to perform the following calibrations:

Optical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
ρ (λ) Spectral Reflectance ^o	Color Values:		White Standard Tile ASTM E1164 D2244
	CIE L*: 0 to 100	0.29 units	
	CIE a*: -26 to 63	0.21 units	
	CIE b*: -26 to 63	0.18 units	
Spectrophotometers Transmittance ^o	τ : 3 %	0.065 % of reading	Neutral Density Filters Holmium Oxide Glass ASTM E275
	τ : 50 %	0.208 % of reading	
	τ : 90 %	0.384 % of reading	
	λ : 340 nm to 750 nm	0.58 nm	
Gloss/Specular Reflectance Angle of Incline ^o	20°	0.18 Gloss Units	Gardner Gloss and Semi-Gloss Standards ASTM D-523
	60°	0.24 Gloss Units	
	85°	0.26 Gloss Units	
Ev Illuminance Chamber ^o	100 lux to 3 000 lux	2 % of reading	Sekonic C-700
Ev Light Color Chamber ^o	2 856 K	26 K	ASTM D1729
Irradiance: Xenon Test Chamber Filter Daylight Q ^o	0.25 W/m ² to 0.68 W/m ² (@340 nm)	5.8 % of reading	ASTM G130 ASTM G155 Radiometer Q-LAB CR20/340/D
Irradiance: Xenon Test Chamber ^o	0.45 W/m ² to 1.5 W/m ² (@420 nm)	5.8 % of reading	ASTM G130 ASTM G155 Radiometer Q-LAB CR20/420
	20 W/m ² to 70 W/m ² (@TUV)	5.8 % of reading	ASTM G130 ASTM G155 Radiometer Q-LAB CR20/TUV
Irradiance: Xenon Test Chamber Filter UV Extended Q/B ^o	0.25 W/m ² to 0.68 W/m ² (@340 nm)	5.8 % of reading	ASTM G155 Radiometer Q-LAB CR20/ 340/ QB
Irradiance: Xenon Test Chamber Filter Window BSL ^o	0.45 W/m ² to 1.5 W/m ² (@420 nm)	5.8 % of reading	ASTM G155 Radiometer Q-LAB CR20/340/BSL
	20 W/m ² to 70 W/m ²	5.8 % of reading	
Irradiance: UVA 340 ^o	0.68 W/m ² /nm to 1.38 W/m ² /nm (@340 nm)	5.8 % of reading	ASTM G130 ASTM G154 Radiometer Q-LAB CR10
Irradiance: Lamps UVA 351 ^o	0.87 W/m ² /nm to 1.55 W/m ² /nm (@340 nm)	5.8 % of reading	



Certificate of Accreditation: Supplement

Quality Instruments Services de México, S.A. de C.V.

Fray Nicolás de Zamora No.73, Col. El Pueblito

Corregidora, Querétaro, México. C.P. 76900

Contact Name: Alvaro Fernando Gomez Prieto Phone: 442-225-4758

Accreditation is granted to the facility to perform the following calibrations:

Optical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Irradiance: Lamps UVB 313 ⁰	0.67 W/m ² /nm to 1.23 W/m ² /nm (@310 nm)	5.8 % of reading	ASTM G130 ASTM G154
Irradiance: Lamps QFS 40 ⁰	0.48 W/m ² /nm to 0.86 W/m ² /nm (@310 nm)	5.8 % of reading	Radiometer Q-LAB CR10

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations. Example: Outside Micrometer^O would mean that the laboratory performs this calibration onsite at the customer's location.
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.